

Certificate of Analysis

Product Compact Dry Swab
ID-No. 1 002 952 / 1 002 953 / 1 402 954
Lot Number 391002
Expiry Date 31.12.2021

| Test | Specifications | Results* |
|--------------|--|-------------|
| Appearance | Clear light yellow solution | Corresponds |
| Aseptic test | No growth of any bacteria. (30°C for 5 days. Visual check) | Corresponds |
| pH value | 6.8 -7.2 | Corresponds |

This lot meets the specifications.

Date: 20.07.2020

Dr.Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.